




AFPS

Animal Feeds Proficiency Testing Scheme

Scheme Description

LGC
Proficiency Testing
1 Chamberhall Business Park
Chamberhall Green
Bury, BL9 0AP
UK.

Telephone: +44 (0) 161 762 2500
Email: ptcustomerservices@lgcgroup.com
Website: www.lgcstandards.com



0001

Record of issue status and modifications

ISSUE	ISSUE DATE	DETAILS	AUTHORISED BY
9	April 2015	Harmonisation of method names for moisture and water in Samples 1, 3 and 9. Addition of sample 10 for <i>Clostridium perfringens</i> and <i>Clostridium</i> species. Addition of analytes <i>E.Coli</i> and Coliforms to sample 7. Added UKAS logo and disclaimers.	M. Whetton T. Noblett
10	Sept 2015	Added new samples for Listeria (sample 11) & E.coli O157 (sample 12) Removed Hard copy Report information.	T.Noblett A.McCarthy
11	Sept 2016	Units changed to % for sample 1 analytes and sample 3 removed. Amended Sample 5 name from Aflatoxin to Mycotoxin	W.Gaunt
12	Sept 2018	Addition of sample 13 fish-feed / aquaculture Updated method details to 'ALL' for microbiology samples and amended Methods paragraph.	W.Gaunt L. Chesters
13	Aug 2019	Energy added to sample 13. Analytes names updated for sample 13. New silage sample added. Sample 12 removed.	W.Gaunt S.Xystouris T.Noblett
14	September 2020	Unit for Energy changed Addition of new sample 15 for enumeration of yeast, mould, coagulase positive Staphylococci and <i>Bacillus cereus</i> and renamed as a contaminants sample. Removed yeast and mould parameters from sample 7 and replaced with lactic acid bacteria, renamed as a probiotic sample. Added total anaerobic count and sulphite-reducing clostridia parameters to sample 10.	S Xystouris C Taylor
15	April 2021	pH added as an analyte to samples 1,9 and 13	L Fielding

Notes:

Where this document has been translated, the English version shall remain the definitive version

Scheme Aims and Organisation

The primary aim of the Animal Feeds Proficiency Testing Scheme (AFPS) is to enable laboratories performing the analysis of animal feeds to monitor their performance and compare it with that of their peers. AFPS also aims to provide information to participants on technical issues and methodologies relating to testing of animal feedstuffs.

The AFPS scheme year operates from January to December. Further information about AFPS, including test material availability, round despatch dates and reporting deadlines, are available on the current AFPS application form.

Test Materials

Details of test materials available in AFPS are given in Appendix A. The test parameters are continually reviewed to ensure they meet the needs of current laboratory testing and regulatory requirements.

Test material batches are tested for homogeneity for at least one test parameter where deemed appropriate. Details of homogeneity tests performed and results are given in the AFPS Scheme Reports.

Some aspects of the scheme, such as test material production, homogeneity testing and stability assessment, can from time to time be subcontracted. When subcontracting occurs, it is placed with a competent subcontractor and LGC is responsible for this work. The planning of the scheme, the evaluation of performance and the authorisation of the final report will never be subcontracted.

Statistical Analysis

Information on the statistics used in AFPS can be found in the General Protocol and in the Scheme Report. Methods for determining assigned values and the values for SDPA used for individual samples are given in Appendix A

Methods

Methods are listed in PORTAL. Please select the most appropriate method from the list. If none of the methods are appropriate, then please report your method as 'Other' and record a brief description in the Comments Section in PORTAL.

Results and Reports

AFPS results are returned through our electronic reporting software, PORTAL, full instructions for which are provided by email.

AFPS reports will be available on the website within 10 working days of round closure. Participants will be emailed a link to the report when it is available.

APPENDIX A - Description of abbreviations used

Assigned Value (AV) - The assigned value may be derived in the following ways:

- From the robust mean (median) of participant results (RMean). This is the median of participant results after the removal of test results that are inappropriate for statistical evaluation, e.g. miscalculations, transpositions and other gross errors. Generally, the assigned value will be set using results from all methods, unless the measurement is considered method-dependant, in which case the assigned value will be set by method as illustrated in the report tables.
For some analytes, where there is a recognised reference method for that type of measurement, this may be used as the assigned value for a particular analyte i.e. it would be applied to results obtained by any method.

Traceability: Assigned values which are derived from the participant results, or a sub-set of the results are not traceable to an international measurement standard. The uncertainty of assigned values derived in this way is estimated from the participant results, according to ISO 13528.

- From a formulation value (Formulation). This denotes the use of an assigned value derived from sample preparation details, where known and exact quantities of analyte have been used to prepare the sample.

Traceability: Assigned values calculated from the formulation of the test sample are traceable, via an unbroken metrological traceability chain, to an international measurement standard. The measurement uncertainty of the assigned value is calculated using the contributions from each calibration in the traceability chain.

- From a qualitative formulation (Qual Form). This applies to qualitative tests where the assigned value is simply based on the presence/absence of the analyte in the test material.

Traceability: Assigned values calculated from the qualitative formulation of the test sample are traceable to a certified reference standard or a microbiological reference strain.

- From expert labs (Expert). The assigned value for the analyte is provided by an 'expert' laboratory.

Traceability: Assigned values provided by an 'expert' laboratory may be traceable to an international measurement standard, according to the laboratory and the method used. The uncertainty of measurement for an assigned value produced in this way will be provided by the laboratory undertaking the analysis. Details of traceability and the associated uncertainty will be provided in the report for the scheme/round.

Range

This indicates the concentration range at which the analyte may be present in the test material.

SDPA

The SDPA represents the 'standard deviation for proficiency assessment' which is used to assess participant performance for the measurement of each analyte. This may be a fixed value (as stated), a percentage (%) of the assigned value or based on the robust standard deviation of the participant measurement results, either across all methods or by method depending on whether the measurement made is method dependent (see assigned value).

Units

This indicates the units used for the assessment of data and in which participants should report their results. For some analytes in some schemes participants may have a choice of which units to report their results, however, the units stipulated in this scheme description are the default units to which any results reported using allowable alternative results will be converted to.

DP

This indicates the number of decimal places to which participants should report their measurement results.

AFPS Scheme Description

Chemistry samples

Sample PT-AF-01
Supplied as:

Nutritional analysis in animal feed
 125g sample of animal feed

Analyte	Method	AV	Range	SDPA	Units	DP
Moisture	Oven drying, Vacuum oven	RMean	All	0.5	%	2
Crude protein	Dumas, Kjeldahl	RMean	All	5% of AV	%	2
Crude fat	Direct extraction, Acid hydrolysis	RMean	All	10% of AV	%	2
Crude ash	AOAC 942-05, EC 152/2009	RMean	All	5% of AV	%	2
Ash insoluble in hydrochloric acid	EC 152/2009	RMean	All	20% of AV	%	3
Sugars	EC 152/2009	RMean	All	20% of AV	%	3
Crude fibre	EC 152/2009, Fibre analyser (e.g. Fiiibretec), Gafta method10:0, ISO 6865	RMean	All	10% of AV	%	3
Starch	Enzymetric, Polarimetric	RMean	All	10% of AV	%	2
ADF	ISO 13906	RMean	All	10% of AV	%	2
NDF	ISO 16472	RMean	All	10% of AV	%	2
PPD (Pepsin protein digestibility)	AOAC 971.09, ISO 6655	RMean	All	10% of AV	% of total protein	2
pH	pH meter	RMean	All	0.10	-	2

Sample PT-AF-02
Supplied as:

Minerals and elements in animal feed
 125g sample of animal feed

Analyte	Method	AV	Range	SDPA	Units	DP
Arsenic	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	3
Cadmium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	mg/kg	3
Calcium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	g/kg	2
Chloride	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	5%	g/kg	2
Chromium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	2
Cobalt	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	3
Copper	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	mg/kg	2
Iron	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	mg/kg	1

AFPS Scheme Description

Analyte	Method	AV	Range	SDPA	Units	DP
Lead	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	3
Magnesium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	g/kg	2
Manganese	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	mg/kg	1
Mercury	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	3
Phosphorus	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	g/kg	2
Potassium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	g/kg	2
Selenium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	3
Sodium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	g/kg	2
Zinc	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	mg/kg	1

Sample PT-AF-05*
Supplied as:

Mycotoxins in animal feed
 125g sample of animal feed

Analyte	Method	AV	Range	SDPA	Units	DP
Aflatoxin B ₁	HPLC, LC-MS, LC-MS/MS, ELISA, Other	RMean	All	Robust SD	µg/kg	2
Aflatoxin B ₂	HPLC, LC-MS, LC-MS/MS, ELISA, Other	RMean	All	Robust SD	µg/kg	2
Aflatoxin G ₁	HPLC, LC-MS, LC-MS/MS, ELISA, Other	RMean	All	Robust SD	µg/kg	2
Aflatoxin G ₂	HPLC, LC-MS, LC-MS/MS, ELISA, Other	RMean	All	Robust SD	µg/kg	2
Total Aflatoxins	HPLC, LC-MS, LC-MS/MS, ELISA, Other	RMean	All	Robust SD	µg/kg	2
Ochratoxin A	HPLC, LC-MS, LC-MS/MS, ELISA, Other	RMean	All	Robust SD	µg/kg	2

***Test material currently not included in LGC's UKAS Scope of Accreditation**

AFPS Scheme Description

Sample PT-AF-08*
Supplied as:

Minerals and trace elements in premix materials
 125g sample of premix

Analyte	Method	AV	Range	SDPA	Units	DP
Arsenic	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	2
Cadmium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	mg/kg	2
Calcium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	g/kg	2
Chloride	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	5%	g/kg	2
Chromium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	2
Cobalt	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	2
Copper	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	mg/kg	2
Iron	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	mg/kg	1
Lead	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	2
Magnesium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	g/kg	2
Manganese	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	mg/kg	1
Mercury	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	2
Phosphorus	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	g/kg	2
Potassium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	g/kg	2
Selenium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	20%	mg/kg	2
Sodium	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	g/kg	2
Zinc	AAS, Colorimetry, ICP-OES, ICP-MS, Other	RMean	All	10%	mg/kg	1

***Test material currently not included in LGC's UKAS Scope of Accreditation**

AFPS Scheme Description

Sample PT-AF-09*
Supplied as:

Nutritional analysis of wet pet food
 125g sample of wet pet food

Analyte	Method	AV	Range	SDPA	Units	DP
Moisture	Oven drying, Vacuum oven	RMean	All	5	%	2
Crude protein	Dumas, Kjeldahl	RMean	All	5%	%	2
Crude fat	Direct extraction, Acid hydrolysis	RMean	All	10%	%	2
Crude ash	Various	RMean	All	5%	%	2
Ash insoluble in hydrochloric acid	Various	RMean	All	20%	%	2
Sugars	Various	RMean	All	20%	%	2
Crude fibre	Various	RMean	All	10%	%	3
Starch	Enzymatic, Polarimetric	RMean	All	10%	%	2
pH	pH meter	RMean	All	0.10	-	2

Sample PT-AF-13*
Supplied as:

Nutritional analysis of fish feed samples
 125g sample of fish feed

Analyte	Method	AV	Range	SDPA	Units	DP
Energy	Calculation	RMean	All	Robust SD	kJ/100g or kcal/100g	2
Moisture	Oven drying, Vacuum oven	RMean	All	Robust SD	%	2
Crude protein	Dumas, Kjeldahl	RMean	All	Robust SD	%	2
Crude fat	Direct extraction, Acid hydrolysis and extraction	RMean	All	Robust SD	%	2
Crude ash	Drying at 500°C Drying at 525°C Drying at 550°C	RMean	All	Robust SD	%	2
Crude fibre	Various	RMean	All	Robust SD	%	3
pH	pH meter	RMean	All	0.10	-	2

***Test material currently not included in LGC's UKAS Scope of Accreditation**

AFPS Scheme Description

Sample PT-AF-14*
Supplied as:

Nutritional analysis of silage
 125g sample of silage

Analyte	Method	AV	Range	SDPA	Units	DP
Moisture	Oven drying, Vacuum oven	RMean	All	Robust SD	%	2
Crude protein	Dumas, Kjeldahl	RMean	All	Robust SD	%	2
Crude fat	Direct extraction, Acid hydrolysis	RMean	All	Robust SD	%	2
Crude ash	Various	RMean	All	Robust SD	%	2
Crude fibre	Various	RMean	All	Robust SD	%	3
Starch	Enzymatic, Polarimetric	RMean	All	Robust SD	%	2

***Test material currently not included in LGC's UKAS Scope of Accreditation**

Microbiological samples**Sample PT-AF-06****Salmonella presence/absence****Supplied as:**

1 x 25g sample of simulated animal feed

Analyte	Method	AV	Range	SDPA	Units	DP
Detection of <i>Salmonella</i> species	ALL	Qual Form	0 to 1,000	N/A	cfu 25g ⁻¹	N/A

Sample PT-AF-07**Microbiological quality indicators and probiotics****Supplied as:**

1 x 10g sample of simulated animal feed

Analyte	Method	AV	Range	SDPA	Units	DP
Total viable count	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0
Enumeration of Enterobacteriaceae	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0
Enumeration of coliforms	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0
Enumeration of <i>Escherichia coli</i>	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0
Enumeration of lactic acid bacteria	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0

AFPS Scheme Description

Sample PT-AF-10

Clostridium perfringens/species

Supplied as:

1 x 10g sample of simulated animal feed

Analyte	Method	AV	Range	SDPA	Units	DP
Total anaerobic count	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0
Enumeration of sulphite-reducing Clostridia	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0
Enumeration of <i>Clostridium perfringens</i>	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0
Enumeration of <i>Clostridium</i> species	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0

Sample PT-AF-11

Listeria monocytogenes/species

Supplied as:

1 x 25g sample of simulated animal feed

Analyte	Method	AV	Range	SDPA	Units	DP
Detection of <i>Listeria</i> species	ALL	Qual Form	0 to 1000	NA	cfu 25g ⁻¹	NA
Detection of <i>L.monocytogenes</i>	ALL	Qual Form	0 to 1000	NA	cfu 25g ⁻¹	NA

AFPS Scheme Description

Sample PT-AF-15

Microbiological Contaminants

Supplied as:

1 x 10g sample of simulated animal feed

Analyte	Method	AV	Range	SDPA	Units	DP
Enumeration of coagulase positive Staphylococci	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0
Enumeration of <i>Bacillus cereus</i>	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0
Enumeration of yeast	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0
Enumeration of mould	ALL	RMean	0 to 100,000	log ₁₀ 0.35	cfu g ⁻¹	0