



SCOPE OF ACCREDITATION TO ISO 17034:2016

LGC STANDARDS
 276 Abby Road
 Manchester, NH 03103
 Chuck Goudreau Phone: 603 622 7660
 E-mail: chuck.goudreau@lgcgroup.com

REFERENCE MATERIAL PRODUCER

Valid to: May 31, 2019

Certificate Number: 2848.02

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this Reference Material Producer for the production of certified reference materials and reference materials of the following category:

Category	Concentration Ranges	Uncertainty	Method	Analytical Technique
Certified Reference Materials				
High Purity Metals – Spectrochemical Solutions				
Single Element Standards	1 to 50 000 µg/mL	< 0.5 % Relative	NIST High Performance ICP-OES	ICP-OES, ICP-MS
Multi Element Standards	1 to 50 000 µg/mL	+/-0.5 % Relative	ICP-OES Aqueous QD, ICP-MS-Aqueous QD	ICP-OES, ICP-MS
Ion Chromatography Standards (Anions and Cations)				
Single-Ion Standards	1 to 10 000 µg/mL	+/- 0.5 % Relative	IC Aqueous QD	IC
Multi-Ion Standards	1 to 10 000 µg/mL	+/- 0.5 % Relative	IC Aqueous QD	IC
Petroleum Products – Fuel and Lubricants				
Acid Number	0.1 to 3.0 mg KOH/g	2 to 30 % Relative	ASTM D664 ASTM D974	Titrimetry
Base Number	6 to 70 mg KOH/g	+/- 4 to 11 % Relative	ASTM D2896 ASTM D4739	Titrimetry
Chlorine (Cl)	1 to 50 000 µg/mL	+/-1 % Relative	ICP-OES-Aqueous QD, XRF QD	ICP-OES, XRF
FT-IR Standard	1 to 100 ABS/cm	+/-10 to 25 % Relative	ASTM E2412	FT-IR

Category	Concentration Ranges	Uncertainty	Method	Analytical Technique
Certified Reference Materials				
Petroleum Products – Fuel and Lubricants (cont)				
Lead (Pb)	0.001 to 5.0 g/gal	+/-1 % Relative	ICP-OES-Aqueous QD, XRF QD	ICP-OES, XRF
Moisture (H ₂ O)	0.1 to 1.0 %	5 % Relative	ASTM D6304	Karl Fischer, Titrimetry
C6.2 Viscosity	0 to 1000 cSt	0.32 % Relative	ASTM D445	Viscometry
Sulfur (S)	1 to 50 000 µg/g	+/-1 % Relative	ASTM D4294	XRF
Single Element Standards	1 to 300 000 µg/g	1 % Relative	ASTM D5185	ICP-OES, XRF
Multi Element Standards	1 to 50 000 µg/g	1 % Relative	ASTM D5185	ICP-OES, XRF
Wear Metals in Oils				
Single Element Standards	1 to 300 000 µg/g	1 % Relative	ASTM D5185	ICP-OES
Multi Element Standards in Petroleum Products	1 to 50 000 µg/g	+/-1 % Relative	ASTM D5185	ICP-OES

Reference Materials, Quality Control Reference Materials				
Petroleum Products - Fuels and Lubricants				
Particle Count Reference Materials	0 to 100 000 particles	N/A	ASTM D7596	Particle Counting





Accredited Reference Material Producer

A2LA has accredited

LGC STANDARDS

Manchester, NH

This accreditation covers the specific materials listed on the agreed upon Scope of Accreditation. This producer meets the requirements of ISO 17034:2016 *General Requirements for the Competence of Reference Material Producers*. This accreditation demonstrates technical competence for a defined scope and the operation of a quality management system.

Presented this 30th day of August 2017.



President and CEO
For the Accreditation Council
Certificate Number 2848.02
Valid to May 31, 2019

For reference materials to which this accreditation applies, please refer to the reference material producer's Scope of Accreditation.