



21 Creek Circle • Boothwyn, PA 19061 USA
Phone 610.447.1555 • Fax 610.447.1577
info@nkhome.com • www.nkhome.com

DECLARATION OF CONFORMITY

EC DECLARATION OF CONFORMITY ACCORDING TO COUNCIL DIRECTIVE 89/336/EEC

Nielsen-Kellerman Company declares that the following products, manufactured by the Nielsen-Kellerman Company at 21 Creek Circle, Boothwyn, PA 19061 USA:

Product Model	Product Description
Kestrel 1000	Pocket Wind Meter <i>Wind Speed Measuring Instrument (Anemometer)</i>
Kestrel 2000	Pocket Wind and Temperature Meter <i>Combined Wind Speed and Temperature Measuring Instrument (Anemometer / Thermometer)</i>
Kestrel 3000	Pocket Weather Meter <i>Combined Wind Speed, Temperature and Relative Humidity Measuring Instrument (Anemometer / Thermometer / Hygrometer)</i>

Are in conformity with the following standards or normative documents:

EMISSIONS:

EN 61000-6-3:2001 and
EN 61326:1998;

Radiated disturbance

- CISPR 22:1997, Class B

IMMUNITY:

EN 61000-6-1:2001 and
EN 61326:1998;

Electrostatic discharge

- IEC 61000-4-2:1995

Radiated RF interference, AM

- IEC 61000-4-3:1995

Power frequency magnetic field

- IEC 61000-4-8:1994

Name of Authorized Person:

Paul Nielsen, Vice-President, Engineering

Signature of Authorized Person:

Test Period:

28th January 2004

Certificate Date:

16th February 2004