

## Supporting Information

# Enhanced immunoassay using a rotating paper platform for quantitative determination of low abundance protein biomarkers

Aboutaleb Sedighi and Ulrich. J. Krull\*


Department of Chemical and Physical Sciences, University of Toronto Mississauga, 3359

Mississauga Road, Mississauga, Ontario, Canada, L5L 1C6. . Email: ulrich.krull@utoronto.ca


<b>Table of content:</b>	<b>Page</b>
<b>Table S1.</b> The oligonucleotide sequences	S2
<b>Figure S1.</b> Optimization of the number of washing cycles after each step in the sandwich immunoassay reaction.	S3
<b>Figure S2.</b> calibration curves of the EXO method	S4
<b>Figure S3:</b> calibration curve of ELISA.	S5

Table S1. The oligonucleotide sequences


Name	Sequence
Molecular beacon (MB)	5'- /SH/-CTGAGCACAGTCCTCAGCGAAA -/Cy3/-3'
Reporter oligonucleotide (R-oligo)	5'- TTTCGCTGAGGACTGTTTTT -3'
Capture oligonucleotide (C-oligo)	5'- /SH/ -AAA AAC AGT CCT CAG CGA AA -3'



**Figure S1.** Optimization of the number of washing cycles after each step in the sandwich immunoassay reaction. The colorimetric signal of the IA zone was measured for the addition of negative control solutions as well as the BNP solution.



**Figure S2.** (A) Signals obtained from the amplification of R-oligos of 1-5000 pM using the EXO method. (B) the linear calibration curve for the R-oligo concentration in the range of 20-5-1000 pM.



**Figure S3.** The calibration curve prepared for determination of IL-6 using ELISA. The dynamic range is 1 - 250 pM.