

Supplemental Material S2. Mean and standard deviation of signal-to-noise ratio (SNR) and baseline amplitude results for submental muscles for both electrodes.

		Experimental patch		Conventional electrodes		Difference	
		Left	Right	Left	Right	Left	Right
SNR	<i>M</i>	20.64	20.31	19.44	19.65	1.19**	0.65**
	<i>SD</i>	4.67	5.19	4.86	5.43	3.49	3.91
Baseline	<i>M</i>	1.33	1.36	1.66	1.7	-0.33**	-0.34**
	<i>SD</i>	2.0	2.06	2.07	2.14	0.51	0.64

Note. ** $p < .05$.