

Supplemental Material S3. Normalized mean amplitude values for 5-ml and 10-ml swallow trials.

		Experimental patch		Conventional electrodes		Difference	
		Left	Right	Left	Right	Left	Right
5 ml	<i>M</i>	11.57	12.03	12.15	13.04	-0.58**	-1.0027
	<i>SD</i>	6.73	7.13	7.01	7.79	3.69	6.31
10 ml	<i>M</i>	11.9	12.83	12.54	13.94	-0.63**	-1.1**
	<i>SD</i>	7.29	7.66	7.13	8.16	4.54	6.84

Note. These values are reported in percentages, as they are normalized. ** $p < .05$.