



Supplemental Figure S7. Network meta-analysis for Leakage rate of each device.

A



C



B

Blom-Singer LP					
	Provox Vega				
0.54 [0.28; 1.03]		Provox 2			
0.48 [0.26; 0.90]	0.90 [0.73; 1.10]		Nijdam		
0.45 [0.26; 0.79]	0.84 [0.57; 1.24]	0.93 [0.67; 1.30]		Groningen LR	
0.42 [0.24; 0.73]	0.78 [0.54; 1.13]	0.87 [0.64; 1.18]	0.93 [0.81; 1.08]		Provox
0.31 [0.18; 0.53]	0.57 [0.39; 0.84]	0.64 [0.46; 0.88]	0.68 [0.59; 0.79]	0.73 [0.66; 0.82]	