

REPORT

REPORT issued by an Accredited Lab

Mikrolab Stockholm AB**Keeplastics AB**

Hans Svensson
Fredriksdalsgatan 16
602 23 Norrköping

Date
2021-07-06

Client number Test number
221216 **K2104353-00**

TEST DATA

Product: Kee seal blue
 Article number: 15212200
 Test identification: 21-224-1
 Test specification: Sterilitytest
 Date of receipt: 2021-06-11
 Date of test: 2021-06-16
 Tested amount: 2 unit
 Analysed according to: See method
 Sampled by: Emma Ejeblad / Hans Svensson

Linje => August
 Negative control: Approved
 Analysis according to current EU pharm 2.6.1 and USP<71>

TEST RESULTS	Method	Result	Unit
Sterility test	M0005	Approved	


COMMENTS

Sollentuna
 Date: 2021-07-06


 Susanna Nilsson

Responsible for the analysis

* After the methodnumber, states that the method is not accredited.


 Linda Storbjörk
 Responsible for the laboratory

The results relate only to the product tested.

Information on measurement uncertainty for quantitative microbiological analyses is available upon request.

Address:
 Mikrolab Stockholm AB
 Kung Hans Väg 3
 192 68 Sollentuna

Phone:
 +46(0)8-297900
 E-mail: sollentuna.lab@alsglobal.com

Board seat:
 Sollentuna
 Corporate-no:
 556794-2171